

MODEL NUMBER PERFORMANCE SPECIFICATION DOC NO. 3224A6 PS3224A6 IEPE ACCELEROMETER REV F, ECN 13890, 12/04/17



• MINIATURE SIZE

EPOXY SEALED CABLE CONNECTION

ΩZ

Weight, Less Cable Mounting Provision

Connector [2] Case Material

Sensing Element

PERFORMANCE

Sensitivity, ±25% [3] [4]
Range F.S. For ± 5 Volts Output
Frequency Response, ±10%
Resonant Frequency
Equivalent Electrical Noise Floor
Amplitude Non-Linearity [5]
Transverse Sensitivity. Max.

ENVIRONMENTAL

Strain Sensitivity @250με

Maximum Vibration
Maximum Shock
Temperature Range

Magnetic Sensitivity at 100 Gauss

ELECTRICAL

Supply Current [6]
Compliance Voltage Range
Output Impedance TYP
Bias Voltage
Discharge Time Constant
Output Signal Polarity for Acceleration Toward Top
Case Grounding

ENGLISH
0.007
[1]
Integral cable
Titanium
Planar Shear Quartz

0.2	mV/g
20,000	g
1.2 to 20,000	Hz
>95	kHz
0.3	g rms
±1%	%F.S
5%	
0.14	g/με

25,000	grms
22,000	±gpk
-60 to +300	°F
Laser Welded, Epoxy sealed Cable	
0.000009	g/Gauss

2 to 20	
18 to +30	
100	
+7 to +9	١
2.0 to 4.0	
Positive	
[7]	

SI		
0.2	gr	
[1]		
Integral cable		
Titanium		
Planar Shear Quartz		

0.020

196200

215600

-51 to 149

Positive

1.2 to 20,000	Hz
>95	kHz
2.94	m/s ² rms
± 1%	%F.S
5%	
1.37	m/s²/με
	т ,
245000	m/s ² rms

aled Cable		Laser Welded, Epoxy sealed Cable	
g/Gauss		0.8829	m/s ² /T
	•'	•	
		000	
	mA	2 to 20	mA
	V	18 to +30	V
	Ω	100	Ω
	VDC	+7 to +9	VDC
	sec	2.0 to 4.0	sec

This family also includes:

This failing also morages.					
Model	Sensitivity (mV/g)	Range (Gpeak)	Resolution (Grms)	Oper. Temp(°F)	TC (sec)
3224A1	9 to 14	500	0.023	-60 to 300	0.9 to 1.5
3224A2	2 to 5	2500	0.03	-60 to 300	0.3 to 1.0
3224A3	5 to 10	1000	0.037	-60 to 300	1.1 to 1.5
3224A4	1 to 2.5	5000	0.08	-60 to 300	0.3 to 1.0
3224A5	9 to 14	500	0.023	-60 to 300	1.1 to 1.5

Refer to the performance specifications of the products in this family for detailed description.

Supplied Accessories:

- 1) Model 6725 installation removal wrench.
- 2) Sample of " Small Petro Mounting Wax".
- 3) Accredited calibration certificate (ISO 17025)


Notes:

mV/ m/s²

± m/s² pk

°C

- [1] Flat mounting surface for adhesive mount.
- [2] 10-32 Jack at end of integral 3FT.cable
- [3] Measured at 100 Hz, 1 g rms per ISA RP 37.2.
- [4] Actual sensitivity is given on a calibration certificate traceable to ISO 17025, supplied with each instrument
- [5] Measured using zero-based best straight-line method.
- [6] Do not apply power to this device without current limiting, 20 mA MAX.
- To do so will destroy the integral IC amplifier.
- [7] Case is grounded to electrical signal/power ground.





21592 Marilla Street, Chatsworth, California 91311 Phone: 818.700.7818 Fax:818.700.7880 www.dytran.com For permission to reprint this content, please contact info@dytran.com